



Device: **INCEPTA ICD F**

Tachy Mode: **Monitor + Therapy**

V-3: 05 May 2012 09:56, VT, V Rate: 165 min⁻¹

Detail

VT Event Onset

Avg V Rate 165 min⁻¹
 Detection Rhythm ID

At V-Detect

Avg V Rate 161 min⁻¹
 Rate Zone VT
 RhythmID Correlated False
 RhythmMatch™
 SRD Met False
 ATP Timeout False


Attempt 1, Burst V ATP


Elapsed Time 00:00:04
 ATP Information
 Number of Bursts 1

Event Ended

00:00:18

EGM displayed at 25mm per second






VT	VT	VT	VT	VT	VT	VT	VT	VP	VP	VP	VP	VP	VP	VP	VP	VP	--	VS
370	368	373	368	373	370	373	375	300	300	300	300	300	300	300	300	300		623

U17%	U21%	U18%	U20%	U23%	U21%	U26%	RID-
V-Epsd							V-Detect



VS	VS	VS	VS	VS	VS	VS	VS	VS	VS
645	638	650	650	655	665	668	688	698	723

LATITUDE™ Patient Management - Event Detail Report



VS
750

VS
763

LATITUDE™ Patient Management - Event Detail Report

