




Episode Detail: *Brady (ID# 102)*



Assessment Legend: Appropriate Indeterminate Inappropriate

ID#	Assessment	Type	Date	Detected hh:mm	Duration hh:mm:ss	Max V. Rate	Median V. Rate
102	<input checked="" type="checkbox"/>	Brady	17-May-2017	09:54	00:00:06	Min = 67 bpm	68 bpm

Other Details

Activity Level: Inactive



Parameter Settings

Arrhythmia Detection: Brady
 Duration: 4 beats
 Rate (Interval): 40 bpm (1500 ms)

Sensing

Sensitivity: 35 μ V
 Blanking: 150 ms

